



IEC compliant

SEER Holter Monitor ECG Cables

Quality ECG cables for your Holter devices play an important role in ensuring that they perform at the high level you have come to expect from GE Healthcare equipment.

SEER™ 12 ECG cables

10-leadwire cables are available for 12-channel Holter ECG recordings. Cables are tested and verified as compatible with the SEER 12 and CardioMem™ CM 3000-12 BT monitors.

SEER 1000 ECG cables

Cables are tested and verified as compatible with SEER 1000 24-hour, 48-hour and 7-day Holter monitors. Offerings are:

- 3-leadwire cables for 3-channel ECG recordings
- 5-leadwire cables for 2-channel ECG recordings
- 7-leadwire cables for 3-channel ECG recordings

SEER Light ECG cables

Cables are tested and verified as compatible with SEER Light Holter monitors. Offerings are:

- 2-channel ECG recordings
- 3-channel ECG recordings

**Our equipment.
Our accessories.
Simply compatible.**

You can trust GE Healthcare accessories and consumables to keep your clinical devices working efficiently.

- Simplified ordering process
- Fast delivery
- One contact point for expert advice
- Quality support at every step

PRODUCT SPECIFICATIONS

SEER 12 and SEER 1000 ECG Cables

Materials

Conductor:	Silver plated alloy
Shield:	Silver plated alloy
Jacket:	Polyurethane, Elastollan (PUR)

Materials composition

Not made with BPA, natural rubber latex, PVC, DEHP or other phthalates.

Operating environment

Temperature:	5° to 45°C (41° to 113°F)
Humidity:	5% to 95%, non-condensing

Storage

Temperature:	-25° to 70°C (-13° to 158°F)
Humidity:	0% to 90%, non-condensing
Pressure:	-700 kPa - 1060 kPa

Disposal considerations

Dispose as clinical waste, in accordance with hospital policy, local guidelines and regulations.

PART LIST

Part Number	Description	Length	Quantity
2104808-001	SEER 12-Channel Holter ECG Cable, Round	130 cm / 51.2 in	1
2104809-001	SEER 12-Channel Holter ECG Cable, Flat	105 cm / 41.3 in	1
2104818-001	SEER 1000 Leadwire set, 3 leadwire, 3 CH	75 cm / 29.5 in	1
2104819-001	SEER 1000 Leadwire set, 3 leadwire, 3 CH	105 cm / 41 in	1
2104810-001	SEER 1000 Leadwire set, 5 leadwire, 2 CH	75 cm / 29.5 in	1
2104812-001	SEER 1000 Leadwire set, 5 leadwire, 2 CH	105 cm / 41 in	1
2104811-001	SEER 1000 Leadwire set, 7 leadwire, 3 CH	75 cm / 29.5 in	1
2104813-001	SEER 1000 Leadwire set, 7 leadwire, 3 CH	105 cm / 41 in	1

Variations in length mean variations in the trunk cable. Lead length in all variants is 45 cm / 17.7 in (overall).

Distributed by:

GE Healthcare Finland Oy



Manufactured by:

GETEMED Medizin- und
Informationstechnik AG
Oderstraße 77
14513 Teltow, Germany

Packaging materials

Primary:	Kraftliner
----------	------------

Cleaning solutions

Ethyl alcohol	70-80%
Isopropyl alcohol	70-80%
Gutaraldehyde	2% (pH 7.5-8.5)
Hydrogen peroxide	1.5%
Sodium hypochlorite	(5% household bleach): Minimum 1:500 dilution (minimum 100 ppm free chlorine) and maximum dilution 1:10

Do not sterilize or autoclave

Compliance

RoHS (Regulation of Hazardous Substances) compliant
International Electrotechnical Commission (IEC) standards
CE Mark

Product compatibility

- SEER 1000
- SEER 12
- CardioMem™ CM 3000-12 BT

PRODUCT SPECIFICATIONS

SEER 12 and SEER 1000 Accessories

Materials

Carrying pouch, SEER 12:	Nylon textile with PVC coat
Fastening shoulder strap, SEER 12:	Chromatic coated rings (2 sides)
Fastening waist belt, SEER 12:	Nylon textile loop backwards
Carrying pouch SEER 1000:	Neoprene
Fastening shoulder strap, SEER 1000:	150 mm poly black C
Fastening waist belt, SEER 1000:	245MMX, 1.5MM 9000D, webbing 3 twill black

Materials composition

Not made with BPA, natural rubber latex, DEHP or other phthalates.

Storage

Temperature:	-25° to 70°C (-13° to 158°F)
Humidity:	0% to 90%, non-condensing
Pressure:	-700 kPa - 1060 kPa

PART LIST

Part Number	Description	Quantity
2014606-297	SEER 12 carrying pouch, 2 battery compartment	1
2104803-001	SEER 12 lead, shoulder straps, waist belt, 12 leads portable pouch	1
2067634-166	SEER 1000 pouch belt clip	1
2067634-073	SEER 1000 pouch with waist belt and shoulder strap	1



Disposal considerations

Dispose as clinical waste, in accordance with hospital policy, local guidelines and regulations.

Packaging materials

Primary: Kraftliner

Cleaning solutions

Clean the pouch by hand using a mild soap solution or machine wash at 60°C (140°F). Dry thoroughly before use, do not spin dry. Disinfect surface with a cloth slightly moistened with 70% alcohol solution.

Compliance

RoHS (Regulation of Hazardous Substances) compliant
CE Mark

Product compatibility

- SEER 1000
- SEER 12

Distributed by:
GE Healthcare Finland Oy



Manufactured by:
GETEMED Medizin- und
Informationstechnik AG
Oderstraße 77
14513 Teltow, Germany

PRODUCT SPECIFICATIONS

SEER Light Holter Accessories

Materials

Cable Cover:	Urethane, Polyvinyl Chloride (PVC)
Cable Yoke:	PVC, Polypropylene (PP)
Cable Shield Cover:	Polyethylene (PE), PVC, Lead (Pb)
Line:	Zinc (Zn), Copper (Cu)
Carry Case Front:	Artificial Leather, Polyester non-woven, Polyurethane
Carry Case Side:	Synthetic rubber, Neoprene
Metal Fittings:	Brass

Materials composition

Not made with BPA, natural rubber latex, DEHP or other phthalates.

Operating environment

Temperature:	5° to 45°C (41° to 113°F)
Humidity:	5% to 90%, non-condensing

Storage

Temperature:	-25° to 70°C (-13° to 158°F)
Humidity:	0% to 90%, non-condensing

PART LIST

Part Number	Description	Quantity
2104827-001	SEER Light, shoulder strap, waist belt, portable pouch	1
2042686-008	SEER Light carry pouch and 3-channel patient cable, 60 cm / 23.7 in	1
2042686-003	SEER Light/Extend carry pouch and 2-channel patient cable, 144 cm/ 57 in	1
2042686-004	SEER Light/Extend carry pouch and 3-channel patient cable, 144 cm/ 57 in	1
2042686-007	SEER Light/Extend carry pouch and short 2-channel patient cable, 62 cm / 24.5 in	1



Disposal considerations

Dispose as clinical waste, in accordance with hospital policy, local guidelines and regulations.

Packaging materials

Primary: Polyethylene

Cleaning solutions

Clean the pouch by hand using a mild soap solution or machine wash at 60°C (140°F). Dry thoroughly before use, do not spin dry. Disinfect surface with a cloth slightly moistened with 70% alcohol solution.

Compliance

RoHS (Regulation of Hazardous Substances) compliant
International Electrotechnical Commission (IEC) standards
CE Mark

Product compatibility

- SEER Light Recorder
- SEER Light Extend Recorder

Product may not be available in all countries and regions. Contact a GE Healthcare Representative for more information. Please visit www.gehealthcare.com/promotional-locations

© 2020 General Electric Company – All rights reserved.

GE Healthcare reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation. Contact your GE Healthcare representative for the most current information. GE and the GE Monogram, are trademarks of General Electric Company. GE Healthcare, a division of General Electric Company. GE, the GE monogram, CardioMem and SEER are trademarks of General Electric Company. GE Medical Systems, Inc., doing business as GE Healthcare.

DOC21046762 Rev 2 12/20

Distributed by:
GE Healthcare



Manufactured by:
Suzuken Company Ltd.
8 Higashikataha-machi,
Higashi-ku, Nagoya, Aichi, JAPAN